



TempTale® Ultra

Serial #: NEFYN00CNV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-01-06 8:52:08 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-01-04 9:20:44 AM
Stop Time: 2025-01-06 8:52:06 PM
Number of Points: 256
Trip Length: 2 day 11 hr 44 min

Summary Data

Low Extreme: -5.4 °C @ 2025-01-04 2:14:44 PM
High Extreme: 8.3 °C @ 2025-01-06 8:36:44 PM
Mean ± Std Deviation: -1.3 °C ± 1.6 °C
Mean Kinetic Temperature: -1.1 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-866.7 °C min	0 d 0 h 0 m (S/E)	2 d 11 h 44 m	1	2025-01-04 9:20:44 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

