



# TempTale® Ultra

Serial #: NJ9YN04Q8V

## Alarm Status

Alarm 

## Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

## Monitor Read

On : 2025-01-08 8:28:02 PM By : TempTale® Ultra  
Note: All Times in GMT

## Recorded Data

First Point: 2025-01-06 4:05:04 PM  
Stop Time: 2025-01-08 8:28:00 PM  
Number of Points: 225  
Trip Length: 2 day 4 hr 30 min

## Summary Data

Low Extreme: -0.8 °C @ 2025-01-06 5:43:04 PM  
High Extreme: 2.6 °C @ 2025-01-06 11:47:04 PM  
Mean ± Std Deviation: -0.4 °C ± 0.8 °C  
Mean Kinetic Temperature: -0.3 °C  
Activation Energy: 83.144 kJ/mol

## Alarm Summary

| Alarm       | Alarms Settings       | Degree Minutes | Threshold (Type)  | Total Time   | Number of Events | Triggered             | Status |
|-------------|-----------------------|----------------|-------------------|--------------|------------------|-----------------------|--------|
| Ideal range | >=-20.0 °C <=-12.0 °C | --             | Unlimited         | 0 d 0 h 0 m  | 0                | --                    | --     |
| Alarm 1     | Over -12.0 °C         | -2639.2 °C min | 0 d 0 h 0 m (S/E) | 2 d 4 h 30 m | 1                | 2025-01-06 4:05:04 PM | Alarm  |
| Alarm 2     | Under -20.0 °C        | 0.0 °C min     | 0 d 0 h 0 m (S/E) | 0 d 0 h 0 m  | 0                | --                    | OK     |

## Alarm Data

## Originator Notes:

