



TempTale® Ultra

Serial #: NEFYN03MJV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-01-08 9:09:47 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-01-06 8:39:03 AM
Stop Time: 2025-01-08 9:09:45 PM
Number of Points: 260
Trip Length: 2 day 12 hr 40 min

Summary Data

Low Extreme: -5.1 °C @ 2025-01-06 2:01:03 PM
High Extreme: 4.0 °C @ 2025-01-08 9:05:03 PM
Mean ± Std Deviation: -1.6 °C ± 1.5 °C
Mean Kinetic Temperature: -1.5 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-1450.8 °C min	0 d 0 h 0 m (S/E)	2 d 12 h 40 m	1	2025-01-06 8:39:03 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

