



TempTale® Ultra

Serial #: NJ9YN04PRV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-01-11 7:21:11 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-01-09 6:14:36 PM
Stop Time: 2025-01-11 7:21:09 PM
Number of Points: 211
Trip Length: 2 day 1 hr 14 min

Summary Data

Low Extreme: -1.7 °C @ 2025-01-09 6:42:36 PM
High Extreme: 2.8 °C @ 2025-01-10 12:46:36 AM
Mean ± Std Deviation: -0.0 °C ± 0.8 °C
Mean Kinetic Temperature: -0.0 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	2546.3 °C min	0 d 0 h 0 m (S/E)	2 d 1 h 14 m	1	2025-01-09 6:14:36 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

