



TempTale® Ultra

Serial #: NJ9YN05XWV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-01-24 8:26:00 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-01-22 11:28:20 AM
Stop Time: 2025-01-24 8:25:58 PM
Number of Points: 245
Trip Length: 2 day 9 hr 10 min

Summary Data

Low Extreme: -7.2 °C @ 2025-01-22 2:16:20 PM
High Extreme: 4.4 °C @ 2025-01-24 8:24:20 PM
Mean ± Std Deviation: -1.8 °C ± 1.9 °C
Mean Kinetic Temperature: -1.6 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	2073.4 °C min	0 d 0 h 0 m (S/E)	2 d 9 h 10 m	1	2025-01-22 11:28:20 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Originator Notes:

