



TempTale® Ultra

Serial #: NJ9YN076GV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-01-31 3:36:45 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-01-29 12:28:17 PM
Stop Time: 2025-01-31 3:36:43 PM
Number of Points: 220
Trip Length: 2 day 3 hr 20 min

Summary Data

Low Extreme: -4.1 °C @ 2025-01-29 3:16:17 PM
High Extreme: 2.9 °C @ 2025-01-30 12:36:17 AM
Mean ± Std Deviation: -0.8 °C ± 1.4 °C
Mean Kinetic Temperature: -0.6 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	1801.2 °C min	0 d 0 h 0 m (S/E)	2 d 3 h 20 m	1	2025-01-29 12:28:17 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

