



TempTale® Ultra

Serial #: NJ9YN07FYV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-02-01 4:18:25 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-01-30 10:37:47 AM
Stop Time: 2025-02-01 4:18:23 PM
Number of Points: 231
Trip Length: 2 day 5 hr 54 min

Summary Data

Low Extreme: -7.9 °C @ 2025-01-30 11:05:47 AM
High Extreme: 4.9 °C @ 2025-01-31 7:59:47 PM
Mean ± Std Deviation: 0.0 °C ± 2.8 °C
Mean Kinetic Temperature: 0.5 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-445.9 °C min	0 d 0 h 0 m (S/E)	2 d 5 h 54 m	1	2025-01-30 10:37:47 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Originator Notes:

