



TempTale® Ultra

Serial #: NJ9YN07FVV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-02-05 2:18:04 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-02-03 6:18:50 AM
Stop Time: 2025-02-05 2:18:02 PM
Number of Points: 240
Trip Length: 2 day 8 hr

Summary Data

Low Extreme: -4.9 °C @ 2025-02-03 2:00:50 PM
High Extreme: -0.2 °C @ 2025-02-03 6:18:50 AM
Mean ± Std Deviation: -2.5 °C ± 1.2 °C
Mean Kinetic Temperature: -2.4 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-788.1 °C min	0 d 0 h 0 m (S/E)	2 d 8 h 0 m	1	2025-02-03 6:18:50 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Originator Notes:

