



TempTale® Ultra

Serial #: NJ9YN05QQV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-02-15 2:25:19 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-02-13 10:18:33 AM
Stop Time: 2025-02-15 2:23:43 PM
Number of Points: 224
Trip Length: 2 day 4 hr 16 min

Summary Data

Low Extreme: -3.4 °C @ 2025-02-13 2:16:33 PM
High Extreme: -0.4 °C @ 2025-02-14 10:00:33 PM
Mean ± Std Deviation: -2.0 °C ± 0.6 °C
Mean Kinetic Temperature: -2.0 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-1469.4 °C min	0 d 0 h 0 m (S/E)	2 d 4 h 16 m	1	2025-02-13 10:18:33 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

