



TempTale® Ultra

Serial #: NKXYN069SV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-03-21 7:03:08 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-03-19 9:18:24 AM
Stop Time: 2025-03-21 7:03:06 PM
Number of Points: 248
Trip Length: 2 day 9 hr 52 min

Summary Data

Low Extreme: -3.7 °C @ 2025-03-19 1:58:24 PM
High Extreme: 0.9 °C @ 2025-03-21 6:56:24 PM
Mean ± Std Deviation: -1.0 °C ± 1.2 °C
Mean Kinetic Temperature: -0.9 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-1145.1 °C min	0 d 0 h 0 m (S/E)	2 d 9 h 52 m	1	2025-03-19 9:18:24 AM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Originator Notes:

