



TempTale® Ultra

Serial #: NEFYN04Y8V

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-03-24 6:42:46 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-03-22 2:32:17 PM
Stop Time: 2025-03-24 6:42:44 PM
Number of Points: 224
Trip Length: 2 day 4 hr 16 min

Summary Data

Low Extreme: -1.2 °C @ 2025-03-22 5:48:17 PM
High Extreme: 1.9 °C @ 2025-03-23 7:14:17 PM
Mean ± Std Deviation: 0.2 °C ± 0.6 °C
Mean Kinetic Temperature: 0.2 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-1041.7 °C min	0 d 0 h 0 m (S/E)	2 d 4 h 16 m	1	2025-03-22 2:32:17 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

