



# TempTale® Ultra

Serial #: NKJYN017HV

## Alarm Status

Alarm 

## Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

## Monitor Read

On : 2025-03-24 7:12:16 PM By : TempTale® Ultra  
Note: All Times in GMT

## Recorded Data

First Point: 2025-03-22 12:14:10 PM  
Stop Time: 2025-03-24 7:12:14 PM  
Number of Points: 236  
Trip Length: 2 day 7 hr 4 min

## Summary Data

Low Extreme: -5.5 °C @ 2025-03-22 2:06:10 PM  
High Extreme: 0.9 °C @ 2025-03-23 2:22:10 PM  
Mean ± Std Deviation: -1.7 °C ± 1.3 °C  
Mean Kinetic Temperature: -1.6 °C  
Activation Energy: 83.144 kJ/mol

## Alarm Summary

## Alarm Data

| Alarm       | Alarms Settings       | Degree Minutes | Threshold (Type)  | Total Time  | Number of Events | Triggered              | Status |
|-------------|-----------------------|----------------|-------------------|-------------|------------------|------------------------|--------|
| Ideal range | >=-20.0 °C <=-12.0 °C | --             | Unlimited         | 0 d 0 h 0 m | 0                | --                     | --     |
| Alarm 1     | Over -12.0 °C         | 1362.5 °C min  | 0 d 0 h 0 m (S/E) | 2 d 7 h 4 m | 1                | 2025-03-22 12:14:10 PM | Alarm  |
| Alarm 2     | Under -20.0 °C        | 0.0 °C min     | 0 d 0 h 0 m (S/E) | 0 d 0 h 0 m | 0                | --                     | OK     |

## Originator Notes:

