



TempTale® Ultra

Serial #: NKXYN069AV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-03-24 7:47:41 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-03-22 1:22:45 PM
Stop Time: 2025-03-24 7:47:39 PM
Number of Points: 234
Trip Length: 2 day 6 hr 36 min

Summary Data

Low Extreme: -6.6 °C @ 2025-03-22 2:18:45 PM
High Extreme: 0.3 °C @ 2025-03-23 2:34:45 PM
Mean ± Std Deviation: -2.0 °C ± 1.2 °C
Mean Kinetic Temperature: -1.9 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm Data

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	13.1 °C min	0 d 0 h 0 m (S/E)	2 d 6 h 36 m	1	2025-03-22 1:22:45 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Originator Notes:

