



TempTale® Ultra

Serial #: NJ9YN03JRV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-03-28 6:21:48 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-03-26 4:09:43 PM
Stop Time: 2025-03-28 6:21:46 PM
Number of Points: 216
Trip Length: 2 day 2 hr 24 min

Summary Data

Low Extreme: -1.1 °C @ 2025-03-26 5:33:43 PM
High Extreme: 1.3 °C @ 2025-03-27 12:47:43 AM
Mean ± Std Deviation: -0.3 °C ± 0.4 °C
Mean Kinetic Temperature: -0.3 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	2492.6 °C min	0 d 0 h 0 m (S/E)	2 d 2 h 24 m	1	2025-03-26 4:09:43 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

