



TempTale® Ultra

Serial #: NKXYN068KV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-03-28 6:16:44 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-03-26 3:52:33 PM
Stop Time: 2025-03-28 6:16:42 PM
Number of Points: 217
Trip Length: 2 day 2 hr 38 min

Summary Data

Low Extreme: -4.4 °C @ 2025-03-26 7:22:33 PM
High Extreme: 3.5 °C @ 2025-03-27 11:08:33 PM
Mean ± Std Deviation: 0.2 °C ± 1.4 °C
Mean Kinetic Temperature: 0.4 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-2133.7 °C min	0 d 0 h 0 m (S/E)	2 d 2 h 38 m	1	2025-03-26 3:52:33 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

