



# TempTale® Ultra

Serial #: PA3YN00MHV

## Alarm Status

Alarm 

## Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

## Monitor Read

On : 2025-10-01 3:37:04 PM By : TempTale® Ultra  
Note: All Times in GMT

## Recorded Data

First Point: 2025-09-29 9:29:53 AM  
Stop Time: 2025-10-01 3:19:11 PM  
Number of Points: 231  
Trip Length: 2 day 5 hr 54 min

## Summary Data

Low Extreme: -4.7 °C @ 2025-09-29 5:11:53 PM  
High Extreme: 2.3 °C @ 2025-09-30 8:57:53 PM  
Mean ± Std Deviation: -1.1 °C ± 1.5 °C  
Mean Kinetic Temperature: -0.9 °C  
Activation Energy: 83.144 kJ/mol

## Alarm Summary

| Alarm       | Alarms Settings       | Degree Minutes | Threshold (Type)  | Total Time   | Number of Events | Triggered             | Status |
|-------------|-----------------------|----------------|-------------------|--------------|------------------|-----------------------|--------|
| Ideal range | >=-20.0 °C <=-12.0 °C | --             | Unlimited         | 0 d 0 h 0 m  | 0                | --                    | --     |
| Alarm 1     | Over -12.0 °C         | 2550.9 °C min  | 0 d 0 h 0 m (S/E) | 2 d 5 h 54 m | 1                | 2025-09-29 9:29:53 AM | Alarm  |
| Alarm 2     | Under -20.0 °C        | 0.0 °C min     | 0 d 0 h 0 m (S/E) | 0 d 0 h 0 m  | 0                | --                    | OK     |

## Alarm Data

## Originator Notes:

