



TempTale® Ultra

Serial #: PA6YN001XV

Alarm Status

Alarm 

Monitor Configuration

Start Up Delay: 30 min Interval: 14 min

Monitor Read

On : 2025-10-01 3:33:28 PM By : TempTale® Ultra
Note: All Times in GMT

Recorded Data

First Point: 2025-09-29 12:41:26 PM
Stop Time: 2025-10-01 3:09:25 PM
Number of Points: 217
Trip Length: 2 day 2 hr 38 min

Summary Data

Low Extreme: -6.7 °C @ 2025-09-29 1:51:26 PM
High Extreme: 4.2 °C @ 2025-09-30 9:07:26 PM
Mean ± Std Deviation: -1.3 °C ± 2.1 °C
Mean Kinetic Temperature: -1.0 °C
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>=-20.0 °C <=-12.0 °C	--	Unlimited	0 d 0 h 0 m	0	--	--
Alarm 1	Over -12.0 °C	-130.1 °C min	0 d 0 h 0 m (S/E)	2 d 2 h 38 m	1	2025-09-29 12:41:26 PM	Alarm
Alarm 2	Under -20.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

